



Taiwan
NO. N2020090227

條碼檢測報告

Bar Code Verification Report

條碼登記公司 Company Name	財團法人中華民國商品條碼策進會 GS1 TAIWAN		公司前置碼 Company Prefix	4710001
公司地址 Address	(105)-臺北市松山區南京東路四段1號13樓 13F., No. 1, Sec. 4, Nanjing E. Rd., Songshan Dist., Taipei City 105, Taiwan (R.O.C.)			
送檢公司 Submission Company	財團法人中華民國商品條碼策進會 GS1 TAIWAN		公司前置碼 Company Prefix	471200099
人眼識別編碼正確性 Human Readable	通過 Pass			
送驗日期 Submission Date	2020-09-22	檢測日期 Verification Date	2020-09-22 本報告效期為3個月This report is valid for 3 months	
樣本名稱 Name of Sample	隱形眼鏡			
符號攜帶資料 Data Encoded	(01)04710001000010(17)201015(10)ABC123			
樣本型態 Type of Sample	印刷樣稿 Printing Sample			
符號型態 Type of Symbol	GS1-128			
檢測標準 Quality Standard	- ISO/IEC 15416:2016 Automatic identification and data capture techniques -- Bar code print quality test specification -- Linear symbols - CNS16026 自動識別及資料擷取技術--條碼列印品質測試規格--線性符號			
檢測結果 Result	本次檢測等級為 B . 該樣本符合國際標準之檢測要求。 This verification grade is B . This sample meets the requirements defined by international standard.			
備註 Note				

註記：

在配合條碼檢測設備下，本檢測報告(以下為本報告)將確保廠商正確使用GS1前置碼及資料內容，並輔助廠商製作條碼的品質管控。本報告無法完全保證條碼可以被掃描，一旦條碼無法被掃描，本報告僅提供造成的原因參考。本報告一旦發出，即無法修改，如本報告造成任何爭議，GS1 Taiwan將保留修改原報告的權利。

重要通知：
本檢測報告(以下為本報告)僅供檢測廠商(依本報告之公司名稱為準)擁有公開使用的權利；非檢測廠商則禁止公開、散佈或重製本報告。假如貴公司收到的檢測報告廠商名稱有誤，請儘速通知GS1 Taiwan。電話：2545-0011

免責聲明：
本檢測報告(以下為本報告)不具有法律效力及不得作為法庭證據，GS1 Taiwan不會涉入或回應任何有關本報告的法律糾紛。GS1 Taiwan盡可能遵照規範及資訊確保本報告的正確性，如仍發生錯誤，GS1 Taiwan亦毋須承擔責任。

Notes (informative localised)

It is the responsibility of the brand owner to ensure the correct use of the GS1 Company Prefix and the correct allocation of the data content. Rejection of products should not necessarily be based only on an out of specification results Bar Code verifiers are measuring devices and are tools that can be used for assisting in quality control. The results are not absolute in that they do not necessarily prove or disprove that the bar code will scan.

Important Note (normative localised)

This Verification Report may contain privileged and confidential information intended only for the use of the addressee named above. If you are not the intended recipient of this report you are hereby notified that any use, dissemination, distribution or reproduction of this message is prohibited. If you

Disclaimer (legal localised)

This report does not constitute evidence for the purpose of any litigation, and GS1 Taiwan will not enter into any discussion, or respond to any correspondence in relation to litigation.

覆核/Approver:陳昭吟 J. Chen

審核/Auditor:陳昭吟 J. Chen

檢驗者/Examiner: 謝晴燄 Yuki Hsieh

Verification Report



FNC101047100010000101720101510ABC123

GS1-128

Print Quality: **3.2/06/660 (B) - Pass (≥1.5)**

Application Validation: **Pass**

Software Version:	0.16.0	Serial Number:	15772
Scan Time:	下午 03:22:51 2020/9/22 (+0800)	Factory Configuration:	2020/9/4
Decode Time:	下午 03:22:54 2020/9/22 (+0800)	User Calibration:	2020/9/21
Computer Name:	ASPIRE-S13	User Name:	User

X Dimension	262.7 μm
Print Gain	15.3 μm, 6% (±33.0%)
Check Character	Pass: 99
Left Quiet Zone	10 Pass
Right Quiet Zone	10 Pass
Height	12.8 mm
Width	75.2 mm

GS1 Healthcare (bedside) (Pass):	(01)04710001000010(17)201015(10)ABC123
-------------------------------------	--

ISO Parameter	Grade	Percentage
Overall	3.2 (B)	
Minimum Reflectance	4.0 (A)	3.0%
Maximum Reflectance		84.0%
Threshold		43.5%
Reflectance Margin		20.7%
Symbol Contrast	4.0 (A)	80.9%
Minimum Edge Contrast	4.0 (A)	50.1%
Modulation	3.2 (B)	61.9%
Defects	4.0 (A)	5.7%
Decodability	3.9 (A)	63.3%
Decode	4.0 (A)	

Print Quality Standard: ISO/IEC 15416
Symbology Specification: ISO/IEC 15417
Verifier Conformance Standard: ISO/IEC 15426-1

Axicon Camera Verifier 0.16.0
Axicon Auto ID Limited www.axicon.com

Application Validation Report

FNC101047100010000101720101510ABC123

GS1-128

Software Version:	0.16.0	Serial Number:	15772
Scan Time:	下午 03:22:51 2020/9/22 (+0800)	Factory Configuration:	2020/9/4
Decode Time:	下午 03:22:54 2020/9/22 (+0800)	User Calibration:	2020/9/21

GS1 Healthcare (bedside) (Pass): (01)04710001000010(17)201015(10)ABC123

Element	Value	Description	Status	Information
(01)	04710001000010	Global Trade Item Number (GTIN) [n2+n14]	Pass	
(17)	2020/10/15	Expiration Date (YYMMDD) [n2+n6]	Pass	AI (17) must be processed together with AI: GTIN or ITIP or (255)
(10)	ABC123	Batch or Lot Number [n2+an..20]	Pass	AI (10) must be processed together with AI: GTIN or ITIP
Grade	3.2 (B)	Grade must be at least 1.5 (C)	Pass	
Aperture	150 µm (06)	Required aperture: 150 µm (06)	Pass	
X dimension	263 µm	Allowed X dimension range: 170 µm to 495 µm	Pass	
Bar Height	12.8 mm	Height of bars should be at least 12.7 mm	N/a	

Axicon Camera Verifier 0.16.0
Axicon Auto ID Limited www.axicon.com



(01)04710001000010
(17)201015
(10)ABC123

CE



2020-10-15

LOT

ABC123



(01) 0 4710001 00001 0 (17) 201015 (10) ABC123